UV-Visible Photoacoustic Spectroscopy for Aerosol Absorption

Al Fischer & Geoffrey D. Smith Chemistry Department University of Georgia al@uga.edu

Background: Aerosols & Climate

Atmospheric aerosols both scatter and absorb incoming solar radiation, thereby cooling and warming Earth, respectively; the balance is one of the largest uncertainties in climate models.

Black carbon (BC) is a strongly absorbing aerosol, and is the second largest contributor to global warming after CO₂; brown carbon (BrC) is another absorbing aerosol, though absorbs primarily in the UV.

Although important for climate models, direct measurements of aerosol absorption remain scarce due to inherent difficulties in measuring it (weak signal, few methods).

The "MultiPAS" Instruments

We have created a modular, multi-wavlength, single-cell photoacoustic system using diode lasers and a multipass cell. The design is adaptable, compact, and portable. We have created three versions:

- [1] **MultiPAS-IV (4 λ):** 406, 532, 662 (+CRD), and 785 nm;
- [2] MultiPAS-UV (3 λ, with 2 UV): 320 (in progress), 375, and 445 (+CRD) nm; and
- [3] MultiPAS-III (4 λ): 422, 532, and 782 nm (built for R. Saleh, UGA CoE).

MultiPAS-IV MP-IV & MP-UV contain a 1- λ cavity ringdown cell for extinction measurements and calibration.

Illuminate around 1400Hz around 1400Hz Perform FFT around 1400Hz **Acoustic Signal**

Current Applications: Ambient Aerosols, Black & Brown Carbon

- 590 - 950 Current work is focused y=2.51x+0.53Preliminary MP-III, MP-IV, & MP-UV on measurements at $R^2 = 0.9100$ spectrum of BC/BrC mix UV wavelengths to aid from combustion of benzene PAS/CRD in retrieving the AAE of (30 second average) BrC and BC. $= B\lambda$ Quantitatively, the AE-33 is high the PAS due to compared to multiple scattering filter-based AE-33 effects in the aethalometer. Future will attempt to develop wavelength-dependent corrections. 2017-03-05 2017-03-06 2017-03-06 00:00 00:00 00:00 00:00 PAS Absorption (Mm⁻¹) λ (nm) Date & Time

The PAS was run side-by-side with an AE-33 7-wavelength aethalometer (Magee Scientific) while sampling dried ambient air in Athens, GA during early Spring 2017, shown above left and middle. The aerosols likely represent well-aged "background" aerosols. The two instruments agree qualitatively, but the AE-33 reads ~2.6 times higher absorption than the PAS (at 660 nm), likely due to known multiple scattering effects in the filter-based aethalometer [e.g. Saturno et al. (2017), AMT] In-depth comparisons to the AE-33, including measurements of the absorption Ångström exponent (AAE), are ongoing with ambient aerosols and lab-made BC and BrC (above right).

Thanks to Dr. Rawad Saleh and Jay Cheng (UGA College of Engineering) for generating BrC/BC samples, providing the aethalometer, and operating the MP-III. This work was funded by National Science Foundation grants AGS-1241621 & AGS-1638307, with travel funding from the UGA Graduate School. Portions of this work have been submitted to Aerosol Science and Technology (in review).

The PAS's low detection limits

(<1 Mm⁻¹) allow it to detect

ambient aerosol even at very

2-minute rolling average is

Qualitatively, the PAS compares

7-wavelength aethalometer run

an

to

aerosol

shown, right.

favorably

simultaneously.

loadings.

AE-33